

1 **Development of a novel duplex lateral flow test for simultaneous**  
2 **detection of casein and  $\beta$ -lactoglobulin in food**

3 Patricia Galan-Malo<sup>a</sup>, Silvia Pellicer<sup>a</sup>, María D. Pérez<sup>b</sup>, Lourdes Sánchez<sup>b</sup>, Pedro Razquin<sup>a</sup> and  
4 Luis Mata<sup>a</sup>

5 <sup>a</sup>ZEULAB, S.L. c/Bari 25 dplo. 50197 Zaragoza (SPAIN)

6 <sup>b</sup>Departamento de Producción Animal y Ciencia de los Alimentos, Facultad de Veterinaria,  
7 Instituto Agroalimentario de Aragón (IA2) (Universidad de Zaragoza-CITA), Miguel Servet, 177,  
8 50013 Zaragoza (SPAIN)

9 **Abstract**

10 Milk by-products such as whey and caseinate are widely used as ingredients or processing aids  
11 in food industry. However, since they could cause allergic reactions they are included in  
12 Allergen Control Plans.  $\beta$ -lactoglobulin is the major whey protein and caseins are main proteins  
13 in milk. Selection of a unique target to analyze the presence of milk in foods could be  
14 insufficient when the source of milk proteins is unknown. A new test based on lateral flow  
15 immunocromatography that combines the simultaneous and independent detection of both  
16 proteins ( $\beta$ -lactoglobulin and casein) in one rapid test was developed. The assay was validated  
17 according to AOAC guidelines being able to detect  $\beta$ -lactoglobulin (0.5 ppm), casein (2 ppm),  
18 whey and powder milk (1-5 ppm). No cross-reactivity was found with a panel of 38 food  
19 commodities. The method is a rapid and suitable tool to identify milk proteins in processed  
20 food, ingredients, and rinsing water.

21 **Keywords**

22 Milk allergens, multiplex, immunochromatography, HACCP

## 23 1. Introduction

24 Milk is widely used as an ingredient in food industry due to its nutritive value and technological  
25 properties. Moreover, milk is a well-balanced nutrient that is introduced in many meals for  
26 children. In order to protect consumers from pathogenic bacteria, milk is treated using  
27 different thermal processes: pasteurisation, in-container-bottle sterilisation or ultra-high  
28 temperature (UHT) treatment (Claeys et al., 2013; EFSA Panel on Dietetic Products, Nutrition  
29 and Allergies, 2014). Moreover, milk can also be processed by evaporation, spray drying,  
30 filtration, hydrolysis and fermentation during the manufacturing of different dairy products,  
31 such as cheese, yogurt, whey concentrate powder or caseinate (Verhoeckx et al., 2015). Some  
32 of them are ready to eat and others are added as ingredients or used as processing aids by  
33 food manufacturers. An example of the last one is the use of caseinates in the manufacturing  
34 of alcoholic beverages (beer, wine) as clarification processing aids. Wines fined with this  
35 product, even after filtration, may trigger adverse reactions in susceptible individuals (EFSA  
36 Panel on Dietetic Products, Nutrition and Allergies, 2014).

37 In children under 3 years milk was identified as one of the main allergenic sources (Fernández-  
38 Rivas, 2009). Milk allergy prevalence in children under 3 years ranged 2-7.5% in Europe, 2.2%  
39 in Canada, 6.5% in Australia and 13.4% in EEUU (EFSA Panel on Dietetic Products, Nutrition and  
40 Allergies, 2014). However, it is reduced when children grow up as a result of development of  
41 clinical tolerance to milk proteins. Symptoms of this allergy can vary from mild, just a skin rash,  
42 to severe reactions with the most serious manifestation of anaphylactic shock (EFSA Panel on  
43 Dietetic Products, Nutrition and Allergies, 2014). Although food labelling and control related to  
44 food allergens is different across the countries, milk is included among the main 8 foods that  
45 should be considered for allergen labelling by the Codex Alimentarius commission (FAO, 2016)  
46 and in the legislation of most countries such as the European Union, USA, Canada, Japan or

47 Australia (Gendel, 2012). In the past years, food industries have done huge efforts to include  
48 allergens in their Hazard Analysis Critical Control Point (HACCP) plans (Gupta et al., 2017).  
49 However, despite such effort, around 10% of reported food alerts in Europe are still due to the  
50 presence of undeclared allergens on food labels (RASFF, 2017). In USA, undeclared food  
51 allergens reached 47% of all commercial foodstuffs in 2013-2014 (Do, Khuda, & Sharma, 2018).  
52 As long as milk is frequently used as an ingredient or a processing aid in food industry, a great  
53 risk of unintentional presence of this allergen exists. Therefore, the development of rapid and  
54 easy-to-use tools can help food industry to control the risk of cross-contaminations as well as  
55 to verify the presence of this ingredient in raw materials and final products.

56 Among the milk proteins, caseins and whey proteins such as  $\beta$ -lactoglobulin or  $\alpha$ -lactalbumin  
57 are the most allergenic. The choice of the most suitable target to detect the presence of milk  
58 proteins depends on several factors, the abundance, the thermal stability and the frequency of  
59 use as ingredient in the manufacturing of foods, among others. Caseins account for 80% of  
60 total proteins in milk; thus respect to abundance, these proteins could be considered as the  
61 best target for milk detection. Meanwhile,  $\beta$ -lactoglobulin is the major whey protein. Caseins  
62 are poorly affected by thermal processes while whey proteins can be denatured, which could  
63 affect their allergenicity (Verhoeckx et al., 2015) and immunodetection (de Luis, Lavilla,  
64 Sánchez, Calvo, & Pérez, 2009). However, casein presence is low in some milk by-products,  
65 such as whey protein isolates or concentrates, which are a very common ingredient in the  
66 manufacturing of many foods. Therefore, detection of milk in food is a challenge and analysing  
67 just one of the two types of proteins as the target could be insufficient to assure the absence  
68 of milk, especially when the source of milk proteins is unknown.

69 Although many methods for detecting food allergens have been developed, only some of them  
70 have a practical application in food industry or a use restricted to some specific conditions.

71 According to FoodDrinkEurope (2013) and Walker (2019), ELISA technique should be used for  
72 mapping the sources of allergen contamination in a facility, to validate the cleaning processes  
73 or to control raw materials and final products since quantitative results may be obtained.  
74 However, for routine cleaning verification checks and for testing finished products on site  
75 Lateral flow immunoassays (LFIA) could be used due to easy to use, quick response and  
76 affordable cost. PCR methods should only be used where no other protein detection  
77 technology is available. Finally, although mass spectrometric methodology is being optimized  
78 for routine allergen food analysis (Monaci, De Angelis, Montemurro, & Pilolli, 2018), it could be  
79 used when secondary confirmatory techniques are required (Walker, 2019).

80 LFIA technique allows a suitable detection of allergenic proteins without the need of trained  
81 operators and well-equipped laboratories, making it a convenient tool to be included in the  
82 HACCP plans (Dzantiev, Byzova, Urusov, & Zherdev, 2014). In addition, due to its quick  
83 response (5-10 min), this method allows *in situ* testing and hence taking corrective actions in a  
84 short time. Up to date, several LFIA tests have been developed to detect casein and/or  $\beta$ -  
85 lactoglobulin and their weaknesses and strengths were recently evaluated (Courtney, Taylor, &  
86 Baumert, 2016). In some cases, test manufacturers claim for detecting total milk proteins in  
87 their products. However, most of them were not able to detect whey proteins but just only  
88 casein. A similar issue was found by using commercial ELISA tests when whey protein  
89 concentrates (WPC) were tested with total milk protein kits (Ivens, Baumert, & Taylor, 2016).  
90 Thus, a great concern is generated when a kit based on the detection of total milk is used to  
91 detect whey ingredients (Courtney et al., 2016; Ivens et al., 2016). The availability of tests  
92 which allow the simultaneous detection of both, casein and whey proteins, could help food  
93 operators to get a complete testing of milk residues in food commodities, independently of the  
94 type of milk proteins added or present by cross-contamination. In addition, the identification  
95 of caseins or  $\beta$ -lactoglobulin in independent lines provides very useful information about the

96 source of contamination, but so far two different tests have to be used to identify the  
97 presence of both milk protein components. In the last years, the development of duplex (or  
98 even multiplex) assays based on lateral flow immunocromatography has increased due to the  
99 advantage of simultaneous detection of several analytes in the same assay. Thus, this format is  
100 excellent for saving time and cost, and for improving control efficiency. Some recent studies  
101 have been focused on mycotoxin detection using multiplex LFIA (Song et al., 2014; Zhang et al.,  
102 2017). However, up to date no multiplex test has been developed to detect allergens in foods  
103 (Ross, Bremer, & Nielen, 2018).

104 The aim of this work has been to develop and evaluate the performance of a novel duplex  
105 lateral flow test for the simultaneous and independent detection of casein and  $\beta$ -lactoglobulin  
106 in a unique LFIA strip. Validation was performed following the AOAC guidelines for qualitative  
107 binary chemistry methods. The POD (Probability Of Detection) was determined at different  
108 levels of the specific target proteins as well as in UHT milk. Cross-reactivity, robustness and lot-  
109 to-lot variation of the test were also evaluated in a single laboratory validation.

## 110 **2. Materials and methods**

### 111 *2.1 Antibody preparation*

112 Affinity purified polyclonal antibodies against  $\beta$ -lactoglobulin were obtained as previously  
113 described (de Luis et al., 2008). Antibodies against  $\beta$ -casein and Internalin A were raised in  
114 rabbits using an approved protocol by the Ethic Committee for Animal Experiments from the  
115 University of Zaragoza (Project Licence PI65/14). Antibodies were purified by affinity  
116 chromatography against the correspondent protein insolubilized in Shepharose (HiTrapNHS-

117 activated HP column, GE) using an AKTAprime plus equipment (GE) and stored at -20 °C until  
118 used.

## 119 *2.2 Preparation of dyed latex particle conjugates for LFIA*

120 Red and Blue Carboxyl-Modified Dyed latex beads (Estapor, Merck) were used as detection  
121 particles. Specific antibodies were coupled to latex beads following manufacturer indications  
122 (EMD Millipore Corporation, 2015). Antibody was added to a final concentration of 0.2 mg mL<sup>-1</sup>  
123 and incubated for 2.5 hours at room temperature with gentle shaking (Rotator, JP Selecta) at  
124 12 rpm and 90° angle. To stop the reaction 30 µL of ethanolamine per mL were added and  
125 incubated for 30 min at room temperature. Then, sample was centrifuged (Sigma 1-16K) at  
126 17.000 x g for 15 min and the supernatant was discarded. For blocking beads, BSA at 1% was  
127 added and incubated for 2 hours at room temperature with gentle shaking. Quality of the  
128 conjugate was assessed by dynamic light scattering (Zetasizer Nano Range, Malvern  
129 Instruments).

130 Conjugates were immediately mixed together at a ratio 1:1:2 (anti-β-lactoglobulin:anti-  
131 casein:anti-Internalin A) and dispensed over the conjugate pad (glass fiber membrane, GE)  
132 with a ZX 1010 Dispenser (Bio-Dot, Irvine, USA).

## 133 *2.3 Preparation of LFIA strips*

134 The anti-casein antibodies for test line 1 (TL1), anti-β-lactoglobulin for test line 2 (TL2) and  
135 recombinant Internalin A for control line (CL), were applied over a nitrocellulose membrane at  
136 1 mg mL<sup>-1</sup> in 3 independent lines using a ZX 1010 dispenser (Bio-Dot, Irvine, USA) as shown in  
137 Figure 1.

138 To assembly the strip components, nitrocellulose membrane, conjugate pad and adsorbent  
139 pads were placed on an adhesive baking card with an overlapping among the components of 2

140 mm to ensure correct capillary flow. Cards were cut into 4 mm-wide strips with a CM4000  
141 Guillotine Cutter (Bio-Dot, Irvine, USA) and stored with desiccant in closed tubes at room  
142 temperature.

#### 143 *2.4 Test procedure*

144 An amount of 1 g or mL of food was mixed with 10 mL of extraction buffer using a filter-plastic  
145 bag (BAGPAGE®, Interscience, France). A volume of 150 µL of filtered sample was placed in a  
146 test tube. Then, the strip was introduced into the sample and incubated for 10 min. A negative  
147 result is obtained when only the control line appears, but none of the test lines. When the  
148 control line and one or both test lines appear, the result is considered positive. If the control  
149 line is not displayed, the result is considered as invalid.

#### 150 *2.5 Validation study*

151 Validation study was performed according to AOAC guidelines. As indicated in Appendix F of  
152 the Guidelines for Standard Method Performance Requirements (AOAC International, 2016),  
153 the performance evaluation for single laboratory validation of qualitative methods have to  
154 include: inclusivity/selectivity, exclusivity/cross-reactivity, environmental interference,  
155 laboratory variance and Probability of Detection (POD). POD is the proportion of positive  
156 analytical outcomes for a qualitative method for a given matrix at a given analyte level. POD is  
157 calculated as the number of positive results divided by the total number of tests at each level  
158 of analyte (AOAC International, 2014). Analyses to obtain POD were performed by 3 different  
159 analysts, in different days and 4 batches of the test were included in the study. Confidence  
160 interval levels, LCL (lower control limit) and UCL (upper control limit) were calculated  
161 according to the method described by Wehling, LaBudde, Brunelle, & Nelson, (2011).

162

163     2.6 *Preparation of samples*

164     2.6.1 *POD determination*

165      $\beta$ -lactoglobulin and sodium caseinate (Sigma-Aldrich) were used to determine the limit of  
166     detection for the specific target test lines. Different concentrations of each protein were  
167     prepared in extraction buffer and the probability of detection was calculated after analyzing at  
168     least 20 independent samples at each level of concentration.

169     Commercial UHT milk with a protein content of 3.1% was also used as a sample to obtain the  
170     limit of detection. The probability of detection at different percentages of UHT milk diluted in  
171     extraction buffer was calculated by analyzing at least 20 independent replicates as previously  
172     described.

173     2.6.2 *Cross-reactivity study*

174     Thirty-eight food commodities were selected following the recommendations of AOAC  
175     guidelines (Abbott et al., 2010). Furthermore, milk from mare, sow, buffalo and camel were  
176     also assayed. A minimum quantity of 100 g was mixed thoroughly with a blender in those  
177     matrices with a heterogeneous composition.

178     2.6.3 *Confirmation of the limit of detection with spiked and incurred foods*

179     Several foods declared as milk-free were selected to confirm the limit of detection: red wine,  
180     herb liquor, soy drink with coffee, Frankfurter sausage, cooked ham, vegetable sauté sauce  
181     and soy and rice infant formulas. A minimum quantity of 100 g was mixed thoroughly with a  
182     blender in those matrices with a heterogeneous composition. Then, samples were spiked with

183 0.05% and 1% of UHT milk, except for liquor that was spiked with 0.2% and 1%, to confirm the  
184 limit of detection for  $\beta$ -lactoglobulin and casein test lines.

185 Thermal processed food (bread, sausage and pâté) incurred with 0.1% of commercial non-fat  
186 spray dry milk (NFSM) were prepared as described previously (de Luis et al., 2008). Two  
187 commercial incurred foods (milk on label), chocolate ice cream and chorizo, were included in  
188 the evaluation.

189 Milk by-products were obtained from the correspondent supplier: WPC 33-Whey Protein  
190 Concentrate 33% protein (ILAS, S.A.), total sodium caseinate (Sigma-Aldrich), skim milk powder  
191 MQA 0902014 (MoniQA Association) and whole milk powder NIST 1549 (National Institute of  
192 Standards and Technology).

193 Sweet whey was prepared from raw milk obtained from a local farm (Movera, Zaragoza,  
194 Spain). Milk was skimmed by centrifuging at 3000xg during 30 min at 4°C. Chymosin from calf  
195 stomach (Sigma-Aldrich) was added to skimmed milk and incubated at 37°C for approximately  
196 1 hour. Then, curd and whey was separated by centrifuging at 3000xg for 30 min at room  
197 temperature. Acid whey was obtained by adding HCl to raw milk slowly until pH was reduced  
198 to 4.6. Whey was separated from precipitated casein by centrifuging at 3000 x g for 30 min.

### 199 **3 Results& Discussion**

#### 200 *3.1 Development of the strip test*

201 To develop a strip test with the ability to detect and identify the two main milk protein  
202 fractions (caseins and whey) two target proteins were selected: casein and  $\beta$ -lactoglobulin.  
203 Thus, three different detection antibodies were combined into the conjugate pad to produce  
204 specific signals in three independent lines, corresponding to control, test 1 for casein and test  
205 2 for  $\beta$ -lactoglobulin.

206 To get an easier interpretation two colored latex particles were combined in the conjugate  
207 pad. Red latex particles were coupled to anti- $\beta$ -lactoglobulin and to anti-casein antibodies to  
208 obtain both test lines whereas blue latex beads were coupled to anti-Internalin A antibodies to  
209 obtain the control line. Different proportions of the three conjugates were tested to obtain the  
210 optimal signal in positive samples without background signal in negative samples. The  
211 proportion 1:1:2 was selected since other conditions showed weak signals in low positive  
212 control samples or unspecific signals with negative control samples.

### 213 *3.2 POD concentration study*

214 The limit of detection of the test was determined independently for each one of the milk  
215 protein targets in the corresponding test lines. The lowest level of detection with a POD value  
216 of 0.95 was 0.5 ppm for  $\beta$ -lactoglobulin (Supplementary material, Table 1) and 2 ppm for  
217 caseinate (Supplementary material, Table 2). UHT milk was also analyzed to determine the  
218 POD on a thermal processed product in which both targets are present (Supplementary  
219 material, Table 3). The UHT treatment was chosen because it is the most common treatment  
220 of liquid milk for consumption. Besides, residues of this product could be found in other liquid  
221 processed foodstuff when manufacturing lines are shared. It could be also a common  
222 contaminant in collective kitchens where it is used for preparing meals instead of milk powder  
223 and other dairy by-products. The lowest level of UHT milk that showed a POD of 1.00 was  
224 0.05%.

225 The typical overloading or hook effect inherent to this kind of tests was also evaluated. This  
226 effect happens when the quantity of the target protein is so high that the binding sites of the  
227 antibodies are saturated resulting in a reduction of the analytical color signal and even giving a  
228 false negative result. The test line for  $\beta$ -lactoglobulin was not displayed when levels of this  
229 protein were higher than 4000 ppm (Supplementary material table 1). When UHT milk was

230 analyzed without additional dilution, the test line for  $\beta$ -lactoglobulin was visualized but the  
231 casein line did not appeared, indicating that the hook effect affected only the later specific  
232 line. Although  $\beta$ -lactoglobulin concentration in milk is around 4 g/L (EFSA Panel on Dietetic  
233 Products, Nutrition and Allergies, 2014), the lack of hook effect could be attributed to the  
234 alteration of some protein epitopes by the heat treatment applied. Our results indicate that  
235 the overloading effect could be compensated with the two test lines, due to the different  
236 content of milk proteins that can be found in food products. Thus, high content of caseins in a  
237 milk powder could produce a hook effect for this milk fraction but not for  $\beta$ -lactoglobulin.  
238 Considering the protein content on the certified reference material of milk powder (MoniQA)  
239 and the  $\beta$ -lactoglobulin content in powder milk (Bobe, Lindberg, Freeman, & Beitz, 2007), only  
240 a concentration higher than 11% (w/w) of milk powder as ingredient in a food sample could  
241 result in a false negative for the  $\beta$ -lactoglobulin test line. In turn, the high content of  $\beta$ -  
242 lactoglobulin in a WPC could produce a hook effect for this protein but not for caseins.

243 Previous studies which compared different food allergen methods for milk protein detection  
244 highlighted an absence of analytical standardization (Johnson et al., 2014; Török et al., 2015).  
245 Following their recommendations, the evaluation of different levels of purified protein targets,  
246 whey and casein fractions, as well as non-fat dry milk (NFDM) reference material was  
247 compared in our study. Table 1 shows the level of detection of casein and  $\beta$ -lactoglobulin in  
248 different dairy by-products widely used in food industry. The test was able to detect 1 ppm of  
249 acid whey and 5 ppm of total sodium caseinate, WPC 33, sweet whey and NFDM. As expected,  
250  $\beta$ -lactoglobulin was poorly detected on sodium caseinate with a level of detection of 100 ppm.  
251 Recently, it was reported that commercial LFIA tests based on the detection of  $\beta$ -lactoglobulin  
252 often fail to detect milk proteins in total sodium caseinate or give a higher level of detection  
253 than the tests based on casein detection (Courtney et al., 2016). Test line based on casein  
254 detection was able to detect low levels of milk protein in WPC. A similar result was also found  
255 with other commercial tests evaluated by Courtney et al. (2016). This is probably due to the

256 fact that some casein molecules are out of the micellar structure and therefore they are  
257 released into whey during curdling, being present in WPC. The LFIA test evaluated in the  
258 present work showed a lower detection limit for  $\beta$ -lactoglobulin than for casein in whey  
259 samples. However, the casein test line was able to detect lower levels of total sodium  
260 caseinate and NFDM than the corresponding line for  $\beta$ -lactoglobulin. In addition, the level of  
261 detection for  $\beta$ -lactoglobulin was different depending on the source of the NFDM material. The  
262 test was able to detect 10 ppm of  $\beta$ -lactoglobulin in MoniQA milk powder reference material  
263 whereas the level of detection was increased to 15 ppm when NIST 1549 milk powder was  
264 used. This result could be due to the different intensity of thermal processing applied to each  
265 product, being the degree of denaturation of  $\beta$ -lactoglobulin higher in NIST than in MoniQA  
266 material. Similar results have been reported by other authors when comparing different egg  
267 powder certified reference materials (Lacorn, Lindeke, Siebeneicher, & Weiss, 2018).

### 268 *3.3 Study of cross-reactivity in food commodities*

269 A panel of 38 food commodities based on AOAC recommendations (Abbott et al., 2010) and  
270 milk from six different animal sources were analyzed using the developed LFIA test. Despite  
271 these food commodities are recommended for ELISA test validation, the same requirement  
272 has been established for LFIA test, as long as both tests are applied for the same purposes and  
273 based on a similar biochemical principle. No cross-reactivity was found for any of the  
274 ingredients analyzed (Table 2). However, in the case of raw meat, seafood and fish the control  
275 line displayed a faint signal, indicating some interference with these samples. Several cheese  
276 and milk samples from different species were also analyzed to determine the cross-reactivity  
277 with bovine proteins. All cheeses made from goat and sheep milk and mozzarella cheese made  
278 with water buffalo milk gave a positive result indicating that the test was able to detect milk  
279 from these species. However, milks from mare, sow and camel gave a negative result with the  
280 test. The reactivity of anti- $\beta$ -lactoglobulin antibodies with sheep, goat and buffalo milk is

281 expected as the protein share about 94-97% of sequence homology in these species. The  
282 absence of reaction with mare and sow milk could be explained by the low homology of  $\beta$ -  
283 lactoglobulin of these species with that of bovine milk (less than 60%). Concerning camel milk,  
284 the lack of reactivity is due to camelids milk is lacking of  $\beta$ -lactoglobulin. In the case of  $\beta$ -  
285 casein, the homology of bovine protein with proteins of sheep, goat and buffalo ranges from  
286 91 to 98% whereas with the protein of mare, sow and camelids this percentage is around 60-  
287 69% (EFSA Panel on Dietetic Products, Nutrition and Allergies, 2014).

#### 288 *3.4 Confirmation of the limit of detection in food matrices (Matrix study)*

289 It is known that the composition of food matrices could hamper the detection of allergens  
290 (Khuda, Jackson, Fu, & Williams, 2015; Poms, Klein, & Anklam, 2004). Therefore, a  
291 representative group of food matrices were spiked at least at two levels of added UHT milk  
292 (0.05 and 1%). UHT milk was selected as the ingredient for spiking since it is a well  
293 standardized product, ease to use and subjected to the most commonly used thermal  
294 processing. The lowest level of milk addition (0.05%) was expected to be only detected with  
295 the casein test line due to the abundance of this protein in milk, whereas the highest level (1%)  
296 would be detected with both, the  $\beta$ -lactoglobulin and casein test lines. The matrices have been  
297 chosen not only following the AOAC recommendations (Abbott et al., 2010), but also  
298 considering features that could challenge the test. As expected, all the matrices analyzed gave  
299 a positive result in both  $\beta$ -lactoglobulin and casein test lines for the level of 1% of UHT milk  
300 added (Table 3). Likewise, a positive result was reached at 0.05% of UHT milk in juice, salad  
301 dressing, soy drink with coffee, sausage, salad dressing, cooked ham, vegetable sauté sauce,  
302 soy infant formula and rice infant formula. However, positive results were only reached at  
303 0.2% of UHT milk in red wine and liquor, indicating that ethanol could produce a slight loss of  
304 sensitivity. It has been reported that ethanol may hamper or weaken the interaction antigen-  
305 antibody (Rehan & Younus, 2006; Singh, Cabello-Villegas, Hutchings, & Mallela, 2010) and thus

306 reduce the ability of antibodies to detect antigens. This fact could be a problem when milk  
307 proteins are used as fining agents to clarify alcoholic beverages. However, sample preparation  
308 includes a 1/10 dilution that would minimize such issue and in addition new procedures could  
309 be adapted for the preparation of this kind of samples. For example, ethanol could be partially  
310 evaporated by a soft heating of the samples preventing from its adverse effect. Nevertheless,  
311 according to EFSA opinion wines fined with casein may trigger adverse reactions in susceptible  
312 individuals and therefore they should be adequately labelled (EFSA Panel on Dietetic Products,  
313 Nutrition and Allergies, 2014; Popping & Diaz-Amigo, 2018). Other samples such as soy drink  
314 with coffee which contains tannins and polyphenols could be detected at 0.05%, indicating  
315 that the extraction buffer was able to neutralize these molecules. This kind of molecules are  
316 well known to interfere with the analysis of food allergens by immunochemical techniques  
317 (Khuda et al., 2015).

### 318 *3.5 Incurred food study*

319 Although spiking a food with milk can be used as a model to explore the matrix effect on the  
320 results of an analysis, incurred samples are preferable as the effect of the industrial processes  
321 on the detection is also evaluated with this kind of samples. Thus, NFDM was added as an  
322 ingredient to several foodstuffs before they were processed. The  $\beta$ -lactoglobulin test line only  
323 appeared at 10 ppm of NFDM in the sausage, but not in bread or pâté (Table 4). By contrast,  
324 the casein was detected at 1 ppm of NFDM in pâté, sausage and bread (Table 4). Pâté and  
325 bread were subjected to a thermal process much more intense than sausage, therefore whey  
326 proteins such as  $\beta$ -lactoglobulin could have undergone a denaturation degree that hampers its  
327 immunodetection. This effect has been previously described by using an ELISA test (de Luis et  
328 al., 2009; Monaci, Brohée, Tregoat, & van Hengel, 2011) and mass spectrometry (Lamberti et  
329 al., 2016). However, caseins are more resistant to such heat treatments and they could be  
330 detected by the LFIA test through the specific line for casein. In a recent study, a LFIA test was

331 developed that could also detect in unique analysis both  $\beta$ -lactoglobulin and casein but  
332 samples of non-dairy incurred foods were not assayed (Masiri et al., 2016). Moreover, that test  
333 was not able to identify the protein source since antibodies against casein and  $\beta$ -lactoglobulin  
334 were mixed in the same test line.

335 The determination of allergens in food using immunoassays has to meet several requirements.  
336 The immunochemical technique needs to be highly sensitive in order to detect a low amount  
337 of hidden allergen. It also has to be highly specific for the target and not to show cross-  
338 reactivity with other food components to avoid false-positive results. Furthermore, another  
339 important issue is related to interferences caused by the matrix components, which may  
340 hinder extraction of the target analyte resulting in underestimation or false-negative results or  
341 which may bind nonspecifically to the antibodies, giving false-positive results (Cucu, Jacxsens,  
342 & De Meulenaer, 2013).

343 The effect of processing is probably the main challenge when using immunoassays to  
344 determine allergens in foods. Processing usually induce physical and chemical modifications of  
345 the target protein that could alter conformational epitopes. Processing may also produce  
346 covalent modifications through Maillard reaction or hydrolysis by fermentation that could alter  
347 lineal epitopes. Furthermore, processing also often cause protein aggregation that may  
348 masked epitopes of the target protein and lose of protein solubility (Gomaa & Boye, 2015;  
349 Monaci et al., 2011).

350 The applied extraction buffers and conditions plays a crucial role in the detection of allergens  
351 in foods, as only successfully extracted target proteins can be detected by their specific  
352 antibodies. The recovery of the extracted target protein depends on the nature of the protein  
353 to be analyzed and the degree of denaturation and aggregation induced by processing  
354 (Steinhoff, Fischer, & Paschke-Kratzin, 2011).

355 On the other hand several studies have shown that results obtained by different commercial  
356 ELISA tests give incomparable quantitative results when they are used to detect allergenic  
357 proteins in processed foods (Gomaa & Boye, 2015; Ivens et al., 2016; Johnson et al., 2014;  
358 Monaci et al., 2011). This variability may be explained in part by the use of different target  
359 proteins, standards and expression of the reported units as well as to the use of different  
360 antibodies and immunoassays formats.

361 At this respect, one of the great concerns in allergen detection is the lack of certified reference  
362 material available. Although recently four testing reference materials for milk allergen has  
363 been validated and commercialized (Poms, 2018), the question of how such materials will be  
364 detected in different real food matrixes subjected to different processing technologies remain  
365 unknown.

### 366 *3.6 Environmental surface testing*

367 Apart from food matrices, the rinsing water after cleaning could be also analyzed by using the  
368 LFIA tests for milk presence in food industry facilities. Thus, some manufacturing plants need  
369 to share production lines for several products including some containing milk or its derivatives.  
370 In these circumstances, an efficient cleaning procedure becomes essential. A useful checkpoint  
371 to control the cleaning efficacy is the rinsing water coming from the Cleaning in Place (CIP)  
372 processes (Jackson et al., 2008). The CIP systems usually alternate acid and basic cleaning  
373 cycles. Hence, the rinse water could contain residues of the acid or basic products that may  
374 affect milk detection. The analysis of water with different concentrations of sodium hydroxide  
375 and acid chloride (0.5, 0.25, 0.1 and 0.05N) showed that no interference occurs at or below  
376 0.1N for both products (Supplementary Material Table 4).

### 377 *3.7 Determination of robustness*

378 The robustness of the method was evaluated by introducing small changes in the standard  
379 procedure both for the extraction and the assay itself. Variations in the sample portion to be  
380 extracted, extraction buffer volume, volume of sample extract to be analyzed and the  
381 temperature of the assay were selected as critical factors. Hence, the effect on sensitivity was  
382 evaluated at two levels of UHT milk addition, 1% for the  $\beta$ -lactoglobulin test line and 0.1% for  
383 the casein test line. The analysis was performed at least in two independent assays for each  
384 condition.

385 To evaluate the effect of the sample portion, three different quantities (0.8, 1.0 and 1.2 g) of  
386 orange juice spiked with UHT milk were mixed with in 10 ml of extraction buffer. Then, the  
387 method was performed following the general instructions. In all conditions, a positive result  
388 was displayed in the  $\beta$ -lactoglobulin test line for the samples spiked with 1% of UHT milk (Table  
389 5). The casein test line gave a positive result for the samples spiked with both 1% and 0.1% of  
390 UHT milk. This result implies that a reduction or increase of 20% in the sample weight neither  
391 affected the limit of detection nor caused matrix interference. The same result was found  
392 when the variation in the volume of extraction buffer was introduced (8, 10 and 12 mL) with  
393 the same sample (Table 5). Thus, a variation of 20% in the volume of the extraction buffer did  
394 not affect the detection limit of both test lines.

395 To evaluate the LFIA method itself, without considering the extraction step, different volumes  
396 (0.1, 0.15 and 0.2 mL) of the sample extract were analyzed. Results are summarized in Table 5.  
397 For all the conditions,  $\beta$ -lactoglobulin test line gave a positive result for the samples spiked  
398 with 1% of UHT milk whereas the casein test line was also positive at 0.1% of added UHT milk.  
399 Different assay times were assayed (2-5-10 min) for the same sample (Table 5). Although both  
400 test lines were positive at 10 min, they were negative after 2 min and only a faint positive  
401 result was shown in the casein test line after 5 min at 0.1% of UHT milk addition. Despite a  
402 positive result was reached after 5 min, to obtain suitable results is important to accomplish

403 the assay until 10 min of incubation, according to the manufacturer indications. In addition,  
404 the juice sample spiked with 1% of UHT milk gave a negative result for the casein test line and  
405 the intensity of the  $\beta$ -lactoglobulin test line was reduced when analyzed at 11°C (result not  
406 shown). Therefore, the results from the robustness evaluation pointed out two critical factors  
407 to obtain a suitable result, the working temperature and the assay time.

### 408 *3.8 Product consistency*

409 The lot-to-lot variation was also evaluated and results are summarized in Table 5 of  
410 supplementary material. The four batches evaluated were able to detect 0.05% of UHT milk  
411 and one reached a level of 0.01%.

## 412 **4 Conclusions**

413 This work shows for the first time a duplex lateral flow immunochromatographic assay able to  
414 detect simultaneously and independently two different allergenic proteins. In addition to alert  
415 for the presence of milk protein residues, the test can identify two of the main allergenic milk  
416 proteins ( $\beta$ -lactoglobulin and  $\beta$ -casein). Because both of these proteins are not always present  
417 in milk ingredients used in food manufacturing or as processing aids, a method able to  
418 simultaneously detect both of them could avoid the choice of only a specific target and would  
419 reduce the number of analysis or analytical runs to be performed by users. Thus, these  
420 advantages simplify the allergen control and save costs. The in-house validation of the new  
421 method has demonstrated a level of detection of 1-5 ppm for whey derivatives through the  $\beta$ -  
422 lactoglobulin test line and 1-5 ppm for caseinate and milk powder with the casein test line.  
423 Moreover, casein test line showed a better level of detection with thermally processed foods.  
424 Method was designed with a blue control line and two red test lines to ease the interpretation  
425 of the results. This assay can be applied for milk detection in raw materials, final processed

426 products or to verify the cleaning procedures. It is suitable to be used as a routine tool in food  
427 industry without the need for specific training of the staff.

#### 428 **Conflicts of interest**

429 The authors declare no conflict of interest exists.

#### 430 **Acknowledgements**

431 This work has been supported by the Ministerio de Economía y Competitividad (Spain) and by  
432 the Fondo Europeo de Desarrollo Regional (European Union). Ref. RTC-2015-3818-2.

#### 433 **References**

- 434 Abbott, M., Hayward, S., Ross, W., Godefroy, S. B., Ulberth, F., Van Hengel, A. J., ... Delahaut, P.  
435 (2010). Validation procedures for quantitative food allergen ELISA methods:  
436 community guidance and best practices. *Journal of AOAC International*, *93*(2), 442–  
437 450.
- 438 AOAC International. (2014). AOAC INTERNATIONAL Guidelines for Validation of Qualitative  
439 Binary Chemistry Methods. *Journal of AOAC International*, *97*(5), 1492–1495.
- 440 AOAC International. (2016). Appendix F: guidelines for standard method performance  
441 requirements. Retrieved from [http://www.eoma.aoc.org/app\\_f.pdf](http://www.eoma.aoc.org/app_f.pdf). Accessed  
442 24/02/2019.
- 443 Bobe, G., Lindberg, G. L., Freeman, A. E., & Beitz, D. C. (2007). Short Communication:  
444 Composition of Milk Protein and Milk Fatty Acids Is Stable for Cows Differing in Genetic  
445 Merit for Milk Production<sup>1</sup>. *Journal of Dairy Science*, *90*(8), 3955–3960.

446 Claeys, W. L., Cardoen, S., Daube, G., De Block, J., Dewettinck, K., Dierick, K., ... Herman, L.  
447 (2013). Raw or heated cow milk consumption: Review of risks and benefits. *Food*  
448 *Control*, 31(1), 251–262.

449 Courtney, R. C., Taylor, S. L., & Baumert, J. L. (2016). Evaluation of Commercial Milk-Specific  
450 Lateral Flow Devices. *Journal of Food Protection*, 79(10), 1767–1774.

451 Cucu, T., Jacxsens, L., & De Meulenaer, B. (2013). Analysis to support allergen risk  
452 management: Which way to go? *Journal of Agricultural and Food Chemistry*, 61(24),  
453 5624–5633.

454 de Luis, R., Lavilla, M., Sánchez, L., Calvo, M., & Pérez, M. D. (2009). Development and  
455 evaluation of two ELISA formats for the detection of  $\beta$ -lactoglobulin in model  
456 processed and commercial foods. *Food Control*, 20(7), 643–647.

457 de Luis, R., Mata, L., Estopañán, G., Lavilla, M., Sánchez, L., & Pérez, M. D. (2008). Evaluation of  
458 indirect competitive and double antibody sandwich ELISA tests to determine  $\beta$ -  
459 lactoglobulin and ovomucoid in model processed foods. *Food and Agricultural*  
460 *Immunology*, 19(4), 339–350.

461 Do, A. B., Khuda, S. E., & Sharma, G. M. (2018). Undeclared Food Allergens and Gluten in  
462 Commercial Food Products Analyzed by ELISA. *Journal of AOAC International*, 101(1),  
463 23–35.

464 Dzantiev, B. B., Byzova, N. A., Urusov, A. E., & Zherdev, A. V. (2014). Immunochromatographic  
465 methods in food analysis. *TrAC Trends in Analytical Chemistry*, 55, 81–93.

466 EFSA Panel on Dietetic Products, Nutrition and Allergies. (2014). Scientific opinion on the  
467 evaluation of allergenic foods and food ingredients for labelling purposes. *EFSA*  
468 *Journal*, 12(11), 3894–4180.

469 EMD Millipore Corporation. (2015). Microsphere coupling - two-step EDC/sulfo NHS covalent  
470 coupling procedure for Estapor Carboxyl-modified dyed microspheres. Billerica, MA  
471 USA. <https://www.sigmaaldrich.com/content/dam/sigma->

472 aldrich/docs/promo\_NOT\_INDEXED/General\_Information/1/microsphere-  
473 coupling.pdf. Accessed 24/02/2019.

474 FAO. (2016). *Handbook on food labelling to protect consumers*. Rome. Retrieved from  
475 <http://www.fao.org/3/a-i6575e.pdf>. Accessed 24/02/2019.

476 Fernández-Rivas, M. (2009). Food allergy in Alergológica-2005. *Journal of Investigational*  
477 *Allergology & Clinical Immunology*, 19 Suppl 2, 37–44.

478 FoodDrinkEurope. (2013). Guidance on food allergen management for food manufactures.  
479 Retrieved from [http://www.fooddrinkeurope.eu/uploads/press-](http://www.fooddrinkeurope.eu/uploads/press-releases_documents/temp_file_FINAL_Allergen_A4_web1.pdf)  
480 [releases\\_documents/temp\\_file\\_FINAL\\_Allergen\\_A4\\_web1.pdf](http://www.fooddrinkeurope.eu/uploads/press-releases_documents/temp_file_FINAL_Allergen_A4_web1.pdf). Accessed 24/02/2019.

481 Gendel, S. M. (2012). Comparison of international food allergen labeling regulations.  
482 *Regulatory Toxicology and Pharmacology*, 63(2), 279–285.

483 Gomaa, A., & Boye, J. (2015). Impact of irradiation and thermal processing on the  
484 immunochemical detection of milk and egg allergens in foods. *Food Research*  
485 *International (Ottawa, Ont.)*, 74, 275–283.

486 Gupta, R. S., Taylor, S. L., Baumert, J. L., Kao, L. M., Schuster, E., & Smith, B. M. (2017).  
487 Economic factors impacting food allergen management: perspectives from the food  
488 industry. *Journal of Food Protection*, 80(10), 1719–1725.

489 Ivens, K. O., Baumert, J. L., & Taylor, S. L. (2016, June 7). Commercial milk enzyme-linked  
490 immunosorbent assay (ELISA) kit reactivities to purified milk proteins and milk-derived  
491 ingredients. *Journal of Food Science*, 81, T1871–T1878.

492 Jackson, L. S., Al-Taher, F. M., Moorman, M., DeVries, J. W., Tippett, R., Swanson, K. M. J., ...  
493 Gendel, S. M. (2008). Cleaning and other control and validation strategies to prevent  
494 allergen cross-contact in food-processing operations. *Journal of Food Protection*, 71(2),  
495 445–458.

496 Johnson, P. E., Rigby, N. M., Dainty, J. R., Mackie, A. R., Immer, U. U., Rogers, A., ... Mills, E. N.  
497 C. (2014). A multi-laboratory evaluation of a clinically-validated incurred quality  
498 control material for analysis of allergens in food. *Food Chemistry*, *148*, 30–36.

499 Khuda, S. E., Jackson, L. S., Fu, T.-J., & Williams, K. M. (2015). Effects of processing on the  
500 recovery of food allergens from a model dark chocolate matrix. *Food Chemistry*, *168*,  
501 580–587.

502 Lacorn, M., Lindeke, S., Siebeneicher, S., & Weiss, T. (2018). Commercial ELISA measurement of  
503 allergens and gluten: what we can learn from case studies. *Journal of AOAC*  
504 *International*, *101*(1), 102–107.

505 Lamberti, C., Acquadro, E., Corpillo, D., Giribaldi, M., Decastelli, L., Garino, C., ... Giuffrida, M.  
506 G. (2016). Validation of a mass spectrometry-based method for milk traces detection  
507 in baked food. *Food Chemistry*, *199*, 119–127.

508 Masiri, J., Barrios-Lopez, B., Benoit, L., Tamayo, J., Day, J., Nadala, C., ... Samadpour, M. (2016).  
509 Development and validation of a lateral flow immunoassay test kit for dual detection  
510 of casein and  $\beta$ -Lactoglobulin residues. *Journal of Food Protection*, *79*(3), 477–483.

511 Monaci, L., Brohée, M., Tregoat, V., & van Hengel, A. (2011). Influence of baking time and  
512 matrix effects on the detection of milk allergens in cookie model food system by ELISA.  
513 *Food Chemistry*, *127*(2), 669–675.

514 Monaci, L., De Angelis, E., Montemurro, N., & Pilolli, R. (2018). Comprehensive overview and  
515 recent advances in proteomics MS based methods for food allergens analysis. *TrAC*  
516 *Trends in Analytical Chemistry*, *106*, 21–36.

517 Poms, R. E. (2018). 2nd MoniQA Symposium. Poster 5. Reference materials for food allergen  
518 analysis. *Quality Assurance and Safety of Crops & Foods*, *10* (Supplement 1), S15.

519 Poms, R. E., Klein, C. L., & Anklam, E. (2004). Methods for allergen analysis in food: a review.  
520 *Food Additives and Contaminants*, *21*(1), 1–31.

521 Popping, B., & Diaz-Amigo, C. (2018). European regulations for labeling requirements for food  
522 allergens and substances causing intolerances: history and future. *Journal of AOAC*  
523 *International*, 101(1), 2–7.

524 RASFF. (2017). The rapid alert system for food and feed, 2016 annual report. Publication Office  
525 of the European Union.  
526 [https://ec.europa.eu/food/sites/food/files/safety/docs/rasff\\_annual\\_report\\_2016.pdf](https://ec.europa.eu/food/sites/food/files/safety/docs/rasff_annual_report_2016.pdf).  
527 Accessed 24/02/2018.

528 Rehan, M., & Younus, H. (2006). Effect of organic solvents on the conformation and interaction  
529 of catalase and anticatalase antibodies. *International Journal of Biological*  
530 *Macromolecules*, 38(3), 289–295.

531 Ross, G. M. S., Bremer, M. G. E. G., & Nielen, M. W. F. (2018). Consumer-friendly food allergen  
532 detection: moving towards smartphone-based immunoassays. *Analytical and*  
533 *Bioanalytical Chemistry*.

534 Singh, S. M., Cabello-Villegas, J., Hutchings, R. L., & Mallela, K. M. G. (2010). Role of partial  
535 protein unfolding in alcohol-induced protein aggregation. *Proteins*, 78(12), 2625–2637.

536 Song, S., Liu, N., Zhao, Z., Njumbe Ediage, E., Wu, S., Sun, C., ... Wu, A. (2014). Multiplex lateral  
537 flow immunoassay for mycotoxin determination. *Analytical Chemistry*, 86(10), 4995–  
538 5001.

539 Steinhoff, M., Fischer, M., & Paschke-Kratzin, A. (2011). Comparison of extraction conditions  
540 for milk and hen's egg allergens. *Food Additives & Contaminants. Part A, Chemistry,*  
541 *Analysis, Control, Exposure & Risk Assessment*, 28(4), 373–383.

542 Török, K., Hajas, L., Horváth, V., Schall, E., Bugyi, Z., Kemény, S., & Tömösközi, S. (2015).  
543 Identification of the factors affecting the analytical results of food allergen ELISA  
544 methods. *European Food Research and Technology*, 241(1), 127–136.

545 Verhoeckx, K. C. M., Vissers, Y. M., Baumert, J. L., Faludi, R., Feys, M., Flanagan, S., ... Kimber, I.  
546 (2015). Food processing and allergenicity. *Food and Chemical Toxicology: An*

547 *International Journal Published for the British Industrial Biological Research*  
548 *Association, 80, 223–240.*

549 Walker, M. J. (2019). Food allergens: an update on analytical methods. In L. Melton, F. Shahidi,  
550 & P. Varelis (Eds.), *Encyclopedia of Food Chemistry* (pp. 622–639). Oxford: Academic  
551 Press.

552 Wehling, P., LaBudde, R. A., Brunelle, S. L., & Nelson, M. T. (2011). Probability of Detection  
553 (POD) as a statistical model for the validation of qualitative methods. *Journal of AOAC*  
554 *International, 94(1), 335–347.*

555 Zhang, X., Yu, X., Wen, K., Li, C., Mujtaba Mari, G., Jiang, H., ... Wang, Z. (2017). Multiplex  
556 lateral flow immunoassays based on amorphous carbon nanoparticles for detecting  
557 three Fusarium mycotoxins in maize. *Journal of Agricultural and Food Chemistry,*  
558 *65(36), 8063–8071.*

559

560 **Figure Captions**

561 Figure 1. Scheme of the distribution of the control and test lines in the strip. CL: Control Line,  
562 TL1: Test line for casein detection, TL2: Test line for  $\beta$ -lactoglobulin detection.

563

### Highlights

A multiplex LFIA test was developed for simultaneous detection of  $\beta$ -lactoglobulin and casein

Detection of  $\beta$ -lactoglobulin and casein was displayed in independent lines

Limit of detection of 0.5 ppm for  $\beta$ -lactoglobulin and 2 ppm for caseins

The new test was evaluated against incurred food processed matrices.



Table 2. Study of cross-reactivity in food commodities. Each sample was analyzed by duplicated. P: positive, P<sup>1</sup>: faint positive, I: invalid, N: negative, LGB:  $\beta$ -lactoglobulin.

Ingredient	Control	Casein	LGB
Almond	P	N	N
Brazil nut	P	N	N
Cashew	P	N	N
Chestnut	P	N	N
Hazelnut	P	N	N
Macadamia nut	P	N	N
Peanut	P	N	N
Pecans	P	N	N
Pine nut	P	N	N
Pistachio	P	N	N
Poppy seed	P	N	N
Pumpkin seed	P	N	N
Sesame	P	N	N
Walnut	P	N	N
Buckwheat	P	N	N
Barley	P	N	N
Corn	P	N	N
Oat	P	N	N
Rye	P	N	N
Wheat	P	N	N
Brown rice	P	N	N
White rice	P	N	N
Chick peas	P	N	N
Lentils	P	N	N
Red beans	P	N	N
Soybean	P	N	N
Split peas	P	N	N
Lupine beans	P	N	N
Cocoa	P	N	N
Coconut	P	N	N
Lecithin	P	N	N
Beef	P <sup>1</sup>	N	N
Chicken	P <sup>1</sup>	N	N
Egg	P	N	N
Pork	P <sup>1</sup>	N	N
Fish	P <sup>1</sup>	N	N
Crustaceans	P <sup>1</sup>	N	N
Gelatin	P	N	N
Camel Milk	P	N	N
Mare Milk	P	N	N
Sow Milk	P	N	N
Buffalo Cheese	P	P	P
Sheep Cheese	P	P	P
Goat Cheese	P	P	P

Table 3. Limit of detection in food matrices spiked with UHT milk. Each sample was analyzed by duplicate in independent assays, except for red wine, liquor and juice which were analyzed by 20 replicates. N: negative, P: Positive, LGB:  $\beta$ -lactoglobulin, POD: Probability of Detection, UCL: upper control limit, LCL: lower control limit, CI: confidence level.

Food Matrix	Spike level	Casein	LGB	POD	LCL (95%CI)	UCL (95%CI)
Orange juice	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N	1.00	0.84	1.00
Red wine	Blank	N	N			
	1% UHT milk	P	P			
	0.2% UHT milk	P	N	1.00	0.84	1.00
Liquor	Blank	N	N			
	1% UHT milk	P	P			
	0.2% UHT milk	P	N	1.00	0.84	1.00
Soy drink with coffee	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N	1.00	0.84	1.00
Sausage	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			
Salad dressing	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			
Cooked ham	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			
Vegetable sauté sauce	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			
Soy infant formula	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			
Rice infant formula	Blank	N	N			
	1% UHT milk	P	P			
	0.05% UHT milk	P	N			

Table 4. Detection of milk in thermal processed foods incurred with non fat spray dried milk. Each sample was analyzed by triplicate. LGB:  $\beta$ -lactoglobulin. N: negative, P: Positive.

Incurred Matrices	Milk level	Casein	LGB
Pâté	Blank	N	N
	100 ppm NFSM	P	N
	10 ppm NFSM	P	N
	5 ppm NFSM	P	N
	1ppm NFSM	P	N
Bread	Blank	N	N
	100 ppm NFSM	P	P
	10 ppm NFSM	P	N
	5 ppm NFSM	P	N
	1ppm NFSM	P	N
Sausage	Blank	N	N
	100 ppm NFSM	P	P
	10 ppm NFSM	P	P
	5 ppm NFSM	P	N
	1ppm NFSM	P	N
Chocolate ice cream (milk on label)		P	P
Chorizo (milk on label)		P	N

Table 5. Results of the robustness study. Effect of the deviations in the protocol on the sensitivity of the test. Each condition was evaluated by duplicate. LGB:  $\beta$ -lactoglobulin. N: negative, P: Positive. P<sup>1</sup>: faint positive

<b>Deviations in the extraction</b>						
Sample portion						
UHT milk (%)	0.8 g		1 g		1.2 g	
	Casein	LGB	Casein	LGB	Casein	LGB
1	P	P	P	P	P	P
0.1	P	N	P	N	P	P <sup>1</sup>
Extraction buffer volume						
UHT milk (%)	8 mL		10 mL		12 mL	
	Casein	LGB	Casein	LGB	Casein	LGB
1	P	P	P	P	P	P
0.1	P	N	P	N	P	N
<b>Deviations in the assay</b>						
Assay volume						
UHT milk (%)	0.1 mL		0.15 mL		0.2 mL	
	Casein	LGB	Casein	LGB	Casein	LGB
1	P	P	P	P	P	P
0.1	P	N	P	N	P	P <sup>1</sup>
Assay time						
UHT milk (%)	2 min		5 min		10 min	
	Casein	LGB	Casein	LGB	Casein	LGB
1	N	N	P	P	P	P
0.1	N	N	P <sup>1</sup>	N	P	N